
Midmark M11 - v1.0.0
Total Cycles: 4437

Name: RITTER 1
Sterilizer ID: V2178198
04/09/2024 08:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 75.9 F 0.2
4:00 123.6 F 0.9
6:00 109.9 F 2.9
8:00 140.1 F 5.6
10:00 206.5 F 9.4
12:00 200.7 F 13.4
14:00 209.9 F 3.8
16:00 224.9 F 6.9
18:00 231.6 F 10.5
20:00 233.2 F 8.3
22:00 238.6 F 10.3
24:00 236.3 F 8.6
26:00 246.4 F 13.2
28:00 257.2 F 19.1
30:00 267.3 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.8 F 28.9
2:00 272.5 F 29.0
3:00 272.1 F 28.9
4:00 271.5 F 28.6

Min 270.9 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.5 F 0.0
5:00 226.1 F 0.0
10:00 224.6 F 0.0
15:00 224.5 F 0.0
20:00 221.5 F 0.0
25:00 233.2 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:19
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:09:36
CYCLE COMPLETE

04/09/2024 09:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____