
Midmark M11 - v1.0.0
Total Cycles: 2707

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2022 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.1 F 0.0
2:00 202.8 F 2.4
4:00 204.9 F 6.7
6:00 213.1 F 12.5
8:00 226.7 F 7.2
10:00 240.4 F 12.8
12:00 238.8 F 10.3
14:00 237.9 F 8.4
16:00 250.9 F 15.6
18:00 262.8 F 22.6
20:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 28.9
2:00 271.8 F 28.6
3:00 272.5 F 28.9
4:00 271.7 F 28.4

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 236.4 F 0.0
10:00 227.8 F 0.0
15:00 229.0 F 0.0
20:00 222.5 F 0.0
25:00 223.5 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 20:10
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

12/21/2022 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____