

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.3uS / 0.8ppm  
CYCLE NUMBER 001467  
15:20 09/30/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:06  
PRESSURIZING 6:55  
STERILIZING 8:28  
136.4 C 220kPa 8:28  
136.3 C 222kPa 8:58  
136.4 C 223kPa 9:28  
136.3 C 222kPa 9:58  
136.3 C 222kPa 10:28  
136.3 C 222kPa 10:58  
136.2 C 222kPa 11:28  
136.4 C 221kPa 11:58  
VENTING 11:58  
AIR DRYING 13:05  
CYCLE COMPLETE 17:53  
-----

Digital Signature #  
6880C387D61688A8  
-----