

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002118  
7:17 26/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:49  
PRESSURIZING 1:56  
STERILIZING 3:07  
136.4 C 220kPa 3:07  
136.3 C 222kPa 3:37  
136.3 C 219kPa 4:07  
135.9 C 221kPa 4:37  
136.3 C 222kPa 5:07  
136.2 C 222kPa 5:37  
136.0 C 221kPa 6:07  
136.1 C 219kPa 6:37  
136.1 C 219kPa 6:37  
VENTING 6:37  
AIR DRYING 7:27  
CYCLE COMPLETE 7:34

-----  
Digital Signature #  
93E7AE3931307862  
-----