

-----  
Midmark M11 - v1.0.5  
Total Cycles: 933

Name: RITTER2  
Sterilizer ID: V2313228  
03/31/2022 06:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.8 F 0.0  
2:00 78.3 F 0.0  
4:00 98.2 F 0.0  
6:00 117.6 F 0.1  
8:00 149.3 F 0.9  
10:00 174.3 F 1.7  
12:00 195.0 F 4.0  
14:00 214.1 F 7.6  
16:00 221.7 F 11.9  
18:00 227.4 F 7.9  
20:00 238.4 F 13.0  
22:00 238.8 F 11.0  
24:00 237.6 F 9.8  
26:00 251.2 F 16.6  
28:00 262.8 F 23.4  
30:00 272.7 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 271.6 F 29.6  
2:00 272.4 F 29.5  
3:00 271.2 F 29.0  
4:00 272.3 F 30.0

Min 271.1 F 28.9  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.3 F 0.0  
5:00 208.7 F 0.0  
10:00 212.0 F 0.0  
15:00 214.4 F 0.0  
20:00 220.2 F 0.0  
25:00 223.8 F 0.0  
30:00 225.4 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 30:10  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:08:14  
CYCLE COMPLETE

03/31/2022 07:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_