
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001105
10:36 03/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:44
STERILIZING 3:54
136.4 C 224kPa 3:54
136.0 C 223kPa 4:24
135.8 C 220kPa 4:54
136.4 C 224kPa 5:24
136.4 C 224kPa 5:54
136.4 C 225kPa 6:24
136.1 C 220kPa 6:54
136.1 C 223kPa 7:24
136.1 C 223kPa 7:24
VENTING 7:24
AIR DRYING 8:13
CYCLE COMPLETE 19:55

Digital Signature #
70F2DCCFA9E1966C
