

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001936  
14:50 08/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:23  
PRESSURIZING 2:34  
STERILIZING 3:28  
136.0 C 221kPa 3:28  
136.4 C 226kPa 3:58  
136.3 C 224kPa 4:28  
136.4 C 225kPa 4:58  
136.3 C 224kPa 5:28  
136.3 C 223kPa 5:58  
136.0 C 222kPa 6:28  
136.1 C 222kPa 6:58  
136.1 C 222kPa 6:58  
VENTING 6:58  
AIR DRYING 7:52  
CYCLE COMPLETE 15:00

-----  
Digital Signature #  
54C3D214C12D8C9B  
-----