
Midmark M11 - v1.0.5
Total Cycles: 2261

Name: RITTER2
Sterilizer ID: V2313228
04/18/2023 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.6 F 0.0
2:00 124.2 F 0.1
4:00 148.1 F 1.2
6:00 178.9 F 5.2
8:00 202.2 F 9.7
10:00 217.6 F 7.8
12:00 226.3 F 9.4
14:00 233.9 F 10.1
16:00 239.4 F 11.8
18:00 238.1 F 10.2
20:00 251.3 F 16.6
22:00 262.8 F 23.3
24:00 272.7 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.5 F 29.8
2:00 271.4 F 29.0
3:00 272.4 F 30.0
4:00 272.1 F 29.5

Min 271.3 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 213.3 F 0.0
10:00 226.7 F 0.0
15:00 226.0 F 0.0
20:00 223.3 F 0.0
25:00 223.2 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:18
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:24
CYCLE COMPLETE

04/18/2023 12:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____