

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1547

Name: RITTER2  
Sterilizer ID: V2313228  
09/29/2022 10:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.0 F 0.0  
2:00 78.1 F 0.0  
4:00 99.2 F 0.0  
6:00 124.5 F 0.2  
8:00 162.9 F 0.3  
10:00 186.2 F 3.0  
12:00 205.3 F 6.8  
14:00 219.1 F 11.0  
16:00 224.8 F 6.0  
18:00 236.0 F 10.6  
20:00 236.1 F 8.3  
22:00 246.3 F 13.1  
24:00 245.4 F 12.7  
26:00 258.0 F 19.5  
28:00 269.2 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.3  
1:00 271.8 F 28.6  
2:00 272.1 F 28.8  
3:00 272.3 F 29.0  
4:00 272.5 F 29.1

Min 271.6 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 233.7 F 0.0  
5:00 211.3 F 0.0  
10:00 220.1 F 0.0  
Dry Cycle Aborted

FILLING: 0:59  
HEATING: 28:54  
STERILIZING: 4:00  
VENTING: 2:44  
DRYING: 11:35  
TOTAL CYCLE: 00:48:12

09/29/2022 11:03 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_