

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 005590  
10:45 24/05/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 1:58  
STERILIZING 2:56  
136.4 C 221kPa 2:56  
136.4 C 224kPa 3:26  
136.4 C 224kPa 3:56  
136.1 C 221kPa 4:26  
136.4 C 225kPa 4:56  
136.2 C 222kPa 5:26  
136.4 C 223kPa 5:56  
136.3 C 220kPa 6:26  
136.3 C 220kPa 6:26  
VENTING 6:26  
AIR DRYING 7:17  
CYCLE COMPLETE 37:17  
-----

Digital Signature #  
4C3DFBCBA50BAA65  
-----