
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000994
10:40 14/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:16
STERILIZING 3:12
136.0 C 228kPa 3:12
136.4 C 233kPa 3:42
136.4 C 233kPa 4:12
136.1 C 231kPa 4:42
136.3 C 233kPa 5:12
136.4 C 231kPa 5:42
136.4 C 233kPa 6:12
136.4 C 233kPa 6:42
136.4 C 233kPa 6:42
VENTING 6:42
AIR DRYING 7:34
CYCLE COMPLETE 11:57

Digital Signature #
F3302580F6126185
