

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1375

Name: RITTER2  
Sterilizer ID: V2313228  
08/08/2022 12:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	115.5 F	0.0
2:00	122.4 F	0.0
4:00	136.2 F	0.1
6:00	173.2 F	1.1
8:00	200.7 F	4.8
10:00	210.5 F	9.2
12:00	226.8 F	11.1
14:00	231.8 F	9.4
16:00	238.1 F	10.4
18:00	242.6 F	12.4
20:00	241.0 F	11.3
22:00	253.5 F	17.6
24:00	264.5 F	24.1

STERILIZING

mm:ss	Degrees	PSI
0:00	271.2 F	28.8
1:00	272.4 F	29.5
2:00	271.7 F	28.8
3:00	271.7 F	29.4
4:00	272.2 F	29.3

Min 271.2 F 28.6  
Max 272.6 F 29.6

VENTING CHAMBER

mm:ss	Degrees	PSI
0:00	207.9 F	0.0
5:00	200.6 F	0.0
10:00	218.4 F	0.0
15:00	214.2 F	0.0
20:00	216.3 F	0.0
25:00	218.6 F	0.0
30:00	218.7 F	0.0

Drying Complete

FILLING: 0:56  
HEATING: 25:60  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:04:07  
CYCLE COMPLETE

08/08/2022 01:22 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_