
Midmark M11 - v1.0.0
Total Cycles: 2799

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2023 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.7 F 0.0
2:00 171.8 F 3.5
4:00 187.2 F 7.3
6:00 216.2 F 12.1
8:00 223.5 F 6.1
10:00 236.8 F 11.3
12:00 234.1 F 8.3
14:00 241.0 F 10.7
16:00 246.2 F 13.3
18:00 258.8 F 20.1
20:00 270.2 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 28.7
2:00 272.1 F 29.0
3:00 272.1 F 28.5
4:00 272.4 F 29.1

Min 271.4 F 28.3
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 237.2 F 0.0
10:00 230.0 F 0.0
15:00 229.1 F 0.0
20:00 227.6 F 0.0
25:00 229.6 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 20:50
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:10
CYCLE COMPLETE

01/19/2023 03:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____