

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1789

Name: RITTER2  
Sterilizer ID: V2313228  
11/30/2022 08:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 79.8 F 0.0  
2:00 82.0 F 0.0  
4:00 85.3 F 0.0  
6:00 105.2 F 0.1  
8:00 146.8 F 0.5  
10:00 178.8 F 0.8  
12:00 203.1 F 3.8  
14:00 215.3 F 7.6  
16:00 227.3 F 11.7  
18:00 229.4 F 7.8  
20:00 241.7 F 13.1  
22:00 240.9 F 11.5  
24:00 239.6 F 10.6  
26:00 253.2 F 17.4  
28:00 264.9 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.9  
1:00 272.6 F 29.6  
2:00 271.8 F 29.2  
3:00 272.1 F 29.1  
4:00 272.4 F 29.7  
  
Min 271.5 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 173.6 F 0.0  
5:00 207.0 F 0.0  
10:00 226.4 F 0.0  
15:00 229.2 F 0.0  
20:00 229.8 F 0.0  
25:00 230.5 F 0.0  
30:00 227.7 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 29:41  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:07:35  
CYCLE COMPLETE

11/30/2022 09:25 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_