

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2497

Name: RITTER 1  
Sterilizer ID: V2178198  
10/26/2022 05:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.1 F 0.0  
2:00 184.4 F 4.4  
4:00 199.3 F 9.3  
6:00 220.4 F 6.3  
8:00 231.6 F 10.4  
10:00 235.9 F 7.3  
12:00 242.1 F 11.2  
14:00 245.7 F 13.4  
16:00 258.7 F 20.4  
18:00 270.3 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.3 F 29.3  
2:00 271.4 F 28.5  
3:00 272.3 F 29.5  
4:00 272.0 F 29.0  
  
Min 271.2 F 28.5  
Max 272.4 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.6 F 0.0  
5:00 228.1 F 0.0  
10:00 239.6 F 0.0  
15:00 237.7 F 0.0  
20:00 229.9 F 0.0  
25:00 226.8 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 1:48  
HEATING: 18:52  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:54  
CYCLE COMPLETE

10/26/2022 06:12 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_