
Midmark M11 - v1.0.5
Total Cycles: 3590

Name: RITTER2
Sterilizer ID: V2313228
05/03/2024 03:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.4 F 0.0
2:00 129.7 F 1.7
4:00 146.7 F 4.8
6:00 173.3 F 8.7
8:00 200.0 F 12.7
10:00 212.7 F 6.4
12:00 226.2 F 10.8
14:00 228.5 F 7.4
16:00 239.1 F 12.2
18:00 238.0 F 10.3
20:00 250.8 F 16.5
22:00 261.7 F 22.8
24:00 272.0 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.9 F 30.0
2:00 271.8 F 29.3
3:00 271.8 F 29.7
4:00 272.6 F 29.9

Min 271.3 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 209.5 F 0.0
10:00 201.2 F 0.0
15:00 199.2 F 0.0
20:00 197.8 F 0.0
25:00 208.3 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 24:35
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:03:14
CYCLE COMPLETE

05/03/2024 04:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____