
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006902
10:38 05/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:44
STERILIZING 4:41
136.1 C 220kPa 4:41
136.1 C 221kPa 5:11
136.5 C 224kPa 5:41
135.9 C 221kPa 6:11
136.4 C 222kPa 6:41
136.0 C 222kPa 7:11
136.3 C 222kPa 7:41
136.1 C 221kPa 8:11
136.1 C 221kPa 8:11
VENTING 8:11
AIR DRYING 9:07
CYCLE COMPLETE 24:15

Digital Signature #
2356367446C90036
