
Midmark M11 - v1.0.5
Total Cycles: 1068

Name: RITTER2
Sterilizer ID: V2313228
05/04/2022 07:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.6 F 0.0
2:00 75.9 F 0.0
4:00 99.0 F 0.0
6:00 104.5 F 0.0
8:00 149.9 F 0.2
10:00 182.4 F 1.0
12:00 196.4 F 4.6
14:00 209.5 F 8.1
16:00 222.3 F 11.8
18:00 226.8 F 6.9
20:00 237.9 F 11.8
22:00 236.3 F 9.4
24:00 239.4 F 9.8
26:00 247.9 F 14.4
28:00 259.4 F 20.9
30:00 270.0 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.5 F 29.3
2:00 271.5 F 28.9
3:00 272.4 F 29.2
4:00 271.5 F 29.0

Min 271.4 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 223.0 F 0.0
10:00 210.3 F 0.0
15:00 199.1 F 0.0
20:00 183.2 F 0.0
25:00 183.0 F 0.0
30:00 180.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:56
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:52
CYCLE COMPLETE

05/04/2022 08:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____