
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000578
10:06 12/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:15
STERILIZING 4:19
136.4 C 230kPa 4:19
136.4 C 232kPa 4:49
136.2 C 231kPa 5:19
136.4 C 233kPa 5:49
136.2 C 230kPa 6:19
136.4 C 233kPa 6:49
136.4 C 233kPa 7:19
136.4 C 233kPa 7:49
136.4 C 233kPa 7:49
VENTING 7:49
AIR DRYING 8:41
CYCLE COMPLETE 23:41

Digital Signature #
727475C1294631E0
