
Midmark M11 - v1.0.5
Total Cycles: 954

Name: RITTER2
Sterilizer ID: V2313228
04/05/2022 06:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.9 F 0.0
2:00 89.6 F 0.0
4:00 92.5 F 0.0
6:00 111.1 F 0.2
8:00 151.1 F 2.9
10:00 171.8 F 6.0
12:00 208.4 F 9.4
14:00 208.0 F 11.8
16:00 225.6 F 6.6
18:00 235.8 F 10.7
20:00 232.8 F 7.8
22:00 243.0 F 12.7
24:00 241.5 F 11.3
26:00 253.6 F 17.4
28:00 264.3 F 23.8
30:00 271.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.8 F 29.4
2:00 271.1 F 28.7
3:00 273.0 F 29.9
4:00 271.9 F 29.3

Min 271.0 F 28.6
Max 273.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.5 F 0.0
5:00 222.2 F 0.0
10:00 199.6 F 0.0
15:00 194.7 F 0.0
20:00 184.0 F 0.0
25:00 192.6 F 0.0
30:00 215.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:06
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:02
CYCLE COMPLETE

04/05/2022 08:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____