
Midmark M11 - v1.0.5
Total Cycles: 2000

Name: RITTER2
Sterilizer ID: V2313228
02/02/2023 12:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.2 F 0.0
2:00 199.8 F 4.0
4:00 200.0 F 9.6
6:00 220.7 F 6.5
8:00 230.6 F 10.6
10:00 232.9 F 8.3
12:00 239.9 F 10.4
14:00 248.1 F 14.8
16:00 260.9 F 22.0
18:00 272.5 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.5 F 29.6
2:00 271.6 F 29.3
3:00 272.4 F 29.6
4:00 271.6 F 29.0

Min 271.5 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 223.1 F 0.0
10:00 243.6 F 0.0
15:00 236.6 F 0.0
20:00 225.4 F 0.0
25:00 222.3 F 0.0
30:00 222.6 F 0.0

Drying Complete

FILLING: 1:44
HEATING: 18:26
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:25
CYCLE COMPLETE

02/02/2023 01:32 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____