
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006966
9:39 22/02/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:03
STERILIZING 3:10
136.5 C 223kPa 3:10
136.2 C 222kPa 3:40
136.4 C 223kPa 4:10
136.4 C 224kPa 4:40
136.3 C 221kPa 5:10
136.1 C 219kPa 5:40
136.4 C 223kPa 6:10
136.1 C 218kPa 6:40
136.1 C 218kPa 6:40
VENTING 6:40
AIR DRYING 7:30
CYCLE COMPLETE 7:35

Digital Signature #
AF3D9D5D30742EE7
