

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4293

Name: RITTER 1  
Sterilizer ID: V2178198  
02/12/2024 12:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.5 F 0.0  
2:00 155.9 F 2.3  
4:00 162.2 F 5.7  
6:00 213.7 F 10.1  
8:00 216.8 F 4.2  
10:00 224.5 F 7.8  
12:00 234.9 F 12.4  
14:00 234.8 F 8.6  
16:00 241.8 F 11.3  
18:00 244.4 F 12.4  
20:00 256.5 F 18.7  
22:00 267.5 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.4 F 28.7  
2:00 271.3 F 28.4  
3:00 272.4 F 28.7  
4:00 271.1 F 28.2

Min 271.0 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.7 F 0.0  
5:00 231.5 F 0.0  
10:00 231.0 F 0.0  
15:00 225.3 F 0.0  
20:00 222.4 F 0.0  
25:00 231.1 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 23:26  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:01:53  
CYCLE COMPLETE

02/12/2024 01:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_