
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000254
12:20 21/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:33
STERILIZING 4:47
136.5 C 224kPa 4:47
136.3 C 224kPa 5:17
136.3 C 225kPa 5:47
136.1 C 222kPa 6:17
136.4 C 221kPa 6:47
136.4 C 225kPa 7:17
136.3 C 224kPa 7:47
136.4 C 225kPa 8:17
136.4 C 225kPa 8:17
VENTING 8:17
AIR DRYING 9:10
CYCLE COMPLETE 9:32

Digital Signature #
49487DCC675F767A
