

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001704  
14:40 07/02/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:04  
STERILIZING 3:07  
136.0 C 222kPa 3:07  
136.3 C 224kPa 3:37  
136.3 C 225kPa 4:07  
136.4 C 225kPa 4:37  
136.4 C 224kPa 5:07  
136.3 C 225kPa 5:37  
136.4 C 226kPa 6:07  
136.0 C 222kPa 6:37  
136.0 C 222kPa 6:37  
VENTING 6:37  
AIR DRYING 7:28  
CYCLE COMPLETE 7:36  
-----

Digital Signature #  
501A3F540DB915DD  
-----