
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000424
12:08 27/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:29
STERILIZING 3:47
136.5 C 234kPa 3:47
136.2 C 230kPa 4:17
136.4 C 231kPa 4:47
136.4 C 233kPa 5:17
136.2 C 232kPa 5:47
136.4 C 233kPa 6:17
136.4 C 233kPa 6:47
136.2 C 231kPa 7:17
136.2 C 231kPa 7:17
VENTING 7:17
AIR DRYING 8:08
CYCLE COMPLETE 8:17

Digital Signature #
0D43CFC3E17873FF
