

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4062

Name: RITTER 1  
Sterilizer ID: V2178198  
11/29/2023 08:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	68.6 F	0.0
2:00	69.0 F	0.2
4:00	80.0 F	1.0
6:00	109.5 F	2.8
8:00	140.4 F	5.0
10:00	164.0 F	7.7
12:00	182.6 F	10.7
14:00	207.9 F	9.0
16:00	210.9 F	4.1
18:00	219.6 F	7.1
20:00	228.2 F	10.4
22:00	232.2 F	9.2
24:00	237.4 F	9.7
26:00	241.2 F	10.6
28:00	243.6 F	12.2
30:00	254.3 F	17.6
32:00	263.8 F	23.3
34:00	272.5 F	28.8

STERILIZING

mm:ss	Degrees	PSI
0:00	272.0 F	28.4
1:00	272.5 F	28.8
2:00	271.7 F	28.8
3:00	271.5 F	28.4
4:00	272.0 F	28.8

Min 271.1 F 28.3  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
0:00	174.2 F	0.0	
5:00	240.3 F	0.0	
10:00	223.5 F	0.0	
15:00	234.1 F	0.0	
20:00	216.8 F	0.0	
25:00	217.4 F	0.0	
30:00	212.9 F	0.0	

Drying Complete

FILLING: 0:51  
HEATING: 34:08  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:12:41  
CYCLE COMPLETE

11/29/2023 09:20 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_