
Midmark M11 - v1.0.0
Total Cycles: 771

Name: RITTER 1
Sterilizer ID: V2178198
04/06/2021 12:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.0 F 0.0
2:00 110.6 F 0.1
4:00 143.4 F 0.1
6:00 150.6 F 1.6
8:00 182.5 F 5.2
10:00 199.8 F 9.2
12:00 221.2 F 12.3
14:00 229.2 F 8.8
16:00 238.8 F 11.3
18:00 241.4 F 11.1
20:00 235.7 F 8.7
22:00 250.2 F 15.4
24:00 262.6 F 22.3
26:00 271.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 28.5
1:00 272.4 F 28.7
2:00 272.6 F 28.9
3:00 272.7 F 29.0
4:00 272.2 F 29.0

Min 271.6 F 28.4
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 211.9 F 0.0
10:00 215.6 F 0.0
15:00 228.9 F 0.0
20:00 226.5 F 0.0
25:00 228.0 F 0.0
30:00 226.4 F 0.0

Drying Complete

FILLING: 0:56
HEATING: 26:41
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:04:34
CYCLE COMPLETE

04/06/2021 02:01 PM
ACCEPTED: X
REJECTED:
Operator: U01