
Midmark M11 - v1.0.0
Total Cycles: 3876

Name: RITTER 1
Sterilizer ID: V2178198
10/05/2023 04:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.6 F 0.0
2:00 179.0 F 4.1
4:00 203.1 F 8.5
6:00 218.1 F 12.6
8:00 229.2 F 8.7
10:00 240.5 F 11.5
12:00 240.9 F 11.7
14:00 239.6 F 10.3
16:00 253.2 F 17.0
18:00 264.7 F 23.7
20:00 271.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.8
1:00 272.1 F 29.1
2:00 272.2 F 29.2
3:00 272.3 F 29.3
4:00 272.0 F 29.2

Min 271.8 F 28.7
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 232.9 F 0.0
10:00 225.9 F 0.0
15:00 224.7 F 0.0
20:00 224.8 F 0.0
25:00 226.8 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

10/05/2023 05:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____