

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2263

Name: RITTER 1  
Sterilizer ID: V2178198  
08/30/2022 08:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 112.0 F 0.0  
2:00 124.8 F 0.0  
4:00 150.1 F 0.2  
6:00 171.4 F 1.1  
8:00 190.8 F 4.0  
10:00 214.5 F 7.9  
12:00 222.0 F 12.4  
14:00 226.6 F 7.2  
16:00 238.2 F 12.1  
18:00 236.3 F 9.2  
20:00 239.4 F 9.8  
22:00 246.7 F 13.6  
24:00 258.5 F 19.9  
26:00 269.1 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.7 F 29.1  
2:00 271.7 F 28.2  
3:00 272.5 F 29.2  
4:00 272.2 F 28.7

Min 271.5 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.5 F 0.0  
5:00 219.4 F 0.0  
10:00 225.7 F 0.0  
15:00 218.2 F 0.0  
20:00 219.8 F 0.0  
25:00 217.6 F 0.0  
Dry Cycle Aborted

FILLING: 1:02  
HEATING: 27:09  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 29:19  
TOTAL CYCLE: 01:04:45

08/30/2022 09:46 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_