
Midmark M11 - v1.0.5
Total Cycles: 2886

Name: RITTER2
Sterilizer ID: V2313228
09/14/2023 03:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.5 F 0.0
2:00 171.8 F 3.4
4:00 192.8 F 7.7
6:00 210.5 F 12.2
8:00 223.0 F 6.6
10:00 229.7 F 11.4
12:00 233.2 F 8.7
14:00 240.1 F 11.1
16:00 246.8 F 13.8
18:00 258.3 F 20.1
20:00 268.7 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.8
1:00 272.1 F 29.5
2:00 272.4 F 29.1
3:00 271.3 F 28.9
4:00 272.5 F 29.5

Min 271.2 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 233.3 F 0.0
10:00 234.1 F 0.0
15:00 228.2 F 0.0
20:00 226.4 F 0.0
25:00 228.4 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:60
HEATING: 21:05
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:00:55
CYCLE COMPLETE

09/14/2023 04:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____