
Midmark M11 - v1.0.0
Total Cycles: 4919

Name: RITTER 1
Sterilizer ID: V2178198
09/18/2024 06:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.0 F 0.0
2:00 171.8 F 2.0
4:00 193.9 F 6.8
6:00 227.0 F 12.6
8:00 225.6 F 8.0
10:00 238.5 F 11.5
12:00 241.0 F 11.4
14:00 238.8 F 9.8
16:00 253.2 F 17.0
18:00 265.7 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.5 F 29.1
2:00 271.7 F 28.3
3:00 272.2 F 29.1
4:00 272.0 F 28.5

Min 271.1 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 253.3 F 0.0
10:00 214.4 F 0.0
15:00 185.7 F 0.0
20:00 185.3 F 0.0
25:00 196.3 F 0.0
30:00 192.0 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 19:30
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

09/18/2024 07:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____