
Midmark M11 - v1.0.0
Total Cycles: 2212

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2022 07:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 80.2 F 0.0
4:00 113.8 F 0.0
6:00 152.6 F 0.6
8:00 175.5 F 2.9
10:00 192.4 F 6.3
12:00 210.6 F 9.9
14:00 220.0 F 11.7
16:00 225.6 F 5.8
18:00 232.4 F 9.7
20:00 242.0 F 11.1
22:00 239.4 F 10.3
24:00 237.5 F 8.2
26:00 247.2 F 14.0
28:00 258.4 F 20.1
30:00 268.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.2
2:00 271.7 F 28.4
3:00 272.4 F 29.1
4:00 272.2 F 28.9

Min 271.5 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 202.9 F 0.0
10:00 215.1 F 0.0
15:00 200.0 F 0.0
20:00 217.6 F 0.0
25:00 225.1 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:14
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:09:09
CYCLE COMPLETE

08/09/2022 08:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____