
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002307
10:40 20/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 2:02
STERILIZING 3:06
136.0 C 220kPa 3:06
136.4 C 226kPa 3:36
136.4 C 225kPa 4:06
136.3 C 221kPa 4:36
136.2 C 225kPa 5:06
136.3 C 224kPa 5:36
136.2 C 223kPa 6:06
136.1 C 223kPa 6:36
136.1 C 223kPa 6:36
VENTING 6:36
AIR DRYING 7:26
CYCLE COMPLETE 19:05

Digital Signature #
8CD79CDB1BBCF835
