
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 005111
16:25 13/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:41
STERILIZING 3:50
136.4 C 222kPa 3:50
136.4 C 225kPa 4:20
136.1 C 222kPa 4:50
136.4 C 225kPa 5:20
136.3 C 223kPa 5:50
136.1 C 219kPa 6:20
136.4 C 222kPa 6:50
136.4 C 223kPa 7:20
136.4 C 223kPa 7:20
VENTING 7:20
AIR DRYING 8:13
CYCLE COMPLETE 25:34

Digital Signature #
5EA4733864FA759A
