
Midmark M11 - v1.0.0
Total Cycles: 802

Name: RITTER 1
Sterilizer ID: V2178198
04/22/2021 02:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.6 F 0.0
2:00 176.7 F 2.6
4:00 187.8 F 5.7
6:00 211.3 F 10.5
8:00 223.0 F 5.3
10:00 232.0 F 9.5
12:00 238.6 F 9.8
14:00 241.1 F 12.0
16:00 239.9 F 10.5
18:00 252.7 F 17.0
20:00 263.9 F 23.6
22:00 271.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.6 F 29.4
2:00 272.0 F 28.9
3:00 271.2 F 28.6
4:00 272.3 F 29.5

Min 271.2 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 220.8 F 0.0
10:00 211.1 F 0.0
15:00 203.9 F 0.0
20:00 210.3 F 0.0
25:00 208.3 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:09
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:01:02
CYCLE COMPLETE

04/22/2021 03:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____