

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5213

Name: RITTER 1  
Sterilizer ID: V2178198  
12/20/2024 04:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.1 F 0.0  
2:00 187.4 F 0.3  
4:00 216.2 F 5.7  
6:00 227.9 F 11.4  
8:00 230.3 F 7.0  
10:00 243.5 F 13.0  
12:00 241.6 F 11.0  
14:00 239.1 F 9.7  
16:00 254.3 F 17.2  
18:00 266.8 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.0  
1:00 272.1 F 28.1  
2:00 272.8 F 28.6  
3:00 272.0 F 28.4  
4:00 272.1 F 28.2  
  
Min 271.5 F 27.9  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.4 F 0.0  
5:00 220.1 F 0.0  
10:00 205.5 F 0.0  
15:00 215.0 F 0.0  
20:00 218.1 F 0.0  
25:00 229.0 F 0.0  
30:00 236.7 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 19:23  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:42  
CYCLE COMPLETE

12/20/2024 05:35 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_