
Midmark M11 - v1.0.5
Total Cycles: 2142

Name: RITTER2
Sterilizer ID: V2313228
03/13/2023 01:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.8 F 0.0
2:00 201.2 F 3.8
4:00 226.8 F 9.2
6:00 230.6 F 7.9
8:00 240.1 F 11.4
10:00 237.8 F 9.6
12:00 236.9 F 9.0
14:00 251.2 F 16.1
16:00 263.6 F 23.4
18:00 271.3 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.6 F 29.2
2:00 271.2 F 28.6
3:00 272.1 F 29.5
4:00 272.1 F 29.0

Min 271.0 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 237.6 F 0.0
10:00 228.2 F 0.0
15:00 230.2 F 0.0
20:00 226.4 F 0.0
25:00 227.3 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 2:18
HEATING: 18:01
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:33
CYCLE COMPLETE

03/13/2023 02:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____