

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4207

Name: RITTER 1  
Sterilizer ID: V2178198  
01/16/2024 09:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.7 F 0.0  
2:00 75.5 F 0.3  
4:00 86.9 F 1.2  
6:00 123.7 F 3.4  
8:00 150.5 F 6.1  
10:00 170.8 F 9.5  
12:00 193.2 F 13.4  
14:00 217.8 F 4.6  
16:00 224.9 F 7.9  
18:00 232.6 F 11.6  
20:00 228.7 F 6.5  
22:00 240.0 F 11.4  
24:00 235.7 F 8.5  
26:00 248.6 F 14.5  
28:00 259.3 F 20.4  
30:00 269.0 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.2 F 28.5  
2:00 272.5 F 28.9  
3:00 271.3 F 28.5  
4:00 272.4 F 28.5

Min 271.1 F 28.2  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.1 F 0.0  
5:00 228.7 F 0.0  
10:00 224.8 F 0.0  
15:00 219.1 F 0.0  
20:00 215.1 F 0.0  
25:00 219.8 F 0.0  
30:00 208.1 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 31:04  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:09:22  
CYCLE COMPLETE

01/16/2024 10:19 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_