

-----  
Midmark M11 - v1.0.5  
Total Cycles: 506

Name: RITTER2  
Sterilizer ID: V2313228  
10/29/2021 02:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.3 F 0.0  
2:00 169.8 F 2.5  
4:00 184.2 F 6.4  
6:00 201.8 F 11.1  
8:00 220.9 F 6.7  
10:00 237.8 F 12.5  
12:00 236.3 F 10.2  
14:00 235.0 F 8.6  
16:00 247.9 F 15.2  
18:00 259.4 F 21.8  
20:00 270.3 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.7  
1:00 272.0 F 30.6  
2:00 271.9 F 30.1  
3:00 272.5 F 30.4  
4:00 271.9 F 30.6

Min 270.9 F 29.6  
Max 273.0 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
5:00 195.7 F 0.0  
10:00 197.0 F 0.0  
15:00 193.5 F 0.0  
20:00 191.8 F 0.0  
25:00 193.1 F 0.0  
30:00 199.6 F 0.0  
Drying Complete

FILLING: 1:48  
HEATING: 20:55  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:56  
CYCLE COMPLETE

10/29/2021 03:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_