

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4659

Name: RITTER 1  
Sterilizer ID: V2178198  
06/19/2024 12:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 108.9 F 0.0  
2:00 111.2 F 1.2  
4:00 134.1 F 3.7  
6:00 159.7 F 6.9  
8:00 184.1 F 10.3  
10:00 210.4 F 5.8  
12:00 218.7 F 5.9  
14:00 228.2 F 9.5  
16:00 237.9 F 13.5  
18:00 236.8 F 9.5  
20:00 240.4 F 10.7  
22:00 244.8 F 12.4  
24:00 255.5 F 18.1  
26:00 265.5 F 24.1  
28:00 271.0 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.1  
1:00 271.9 F 28.7  
2:00 272.7 F 28.7  
3:00 271.8 F 28.8  
4:00 272.5 F 28.8

Min 270.9 F 28.0  
Max 273.0 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
5:00 207.3 F 0.0  
10:00 207.8 F 0.0  
15:00 188.3 F 0.0  
20:00 197.6 F 0.0  
25:00 192.8 F 0.0  
30:00 186.3 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 28:02  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:06:49  
CYCLE COMPLETE

06/19/2024 02:05 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_