

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2391

Name: RITTER2  
Sterilizer ID: V2313228  
05/18/2023 03:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.4 F 0.0  
2:00 185.6 F 3.6  
4:00 204.8 F 9.1  
6:00 222.6 F 8.0  
8:00 230.7 F 10.9  
10:00 234.1 F 8.7  
12:00 238.9 F 10.0  
14:00 248.4 F 14.8  
16:00 260.6 F 21.7  
18:00 271.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.0  
1:00 272.4 F 29.7  
2:00 272.0 F 29.1  
3:00 271.3 F 29.1  
4:00 272.3 F 29.8  
  
Min 271.1 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.2 F 0.0  
5:00 233.3 F 0.0  
10:00 235.3 F 0.0  
15:00 228.1 F 0.0  
20:00 228.4 F 0.0  
25:00 231.1 F 0.0  
30:00 228.1 F 0.0  
Drying Complete

FILLING: 2:08  
HEATING: 18:39  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:05  
CYCLE COMPLETE

05/18/2023 04:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_