
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001486
11:05 26/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:14
STERILIZING 4:20
136.4 C 224kPa 4:20
136.1 C 224kPa 4:50
136.4 C 226kPa 5:20
136.3 C 225kPa 5:50
136.4 C 225kPa 6:20
136.3 C 225kPa 6:50
136.4 C 226kPa 7:20
136.2 C 224kPa 7:50
136.2 C 224kPa 7:50
VENTING 7:50
AIR DRYING 8:41
CYCLE COMPLETE 8:51

Digital Signature #
9FB33BA10871C464
