
Midmark M11 - v1.0.0
Total Cycles: 2262

Name: RITTER 1
Sterilizer ID: V2178198
08/30/2022 06:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.4 F 0.0
2:00 93.2 F 0.0
4:00 85.2 F 0.6
6:00 119.8 F 2.8
8:00 159.3 F 5.6
10:00 172.8 F 9.0
12:00 197.8 F 12.7
14:00 212.4 F 3.9
16:00 223.1 F 7.4
18:00 233.9 F 11.1
20:00 231.5 F 6.7
22:00 241.1 F 11.5
24:00 236.7 F 8.9
26:00 249.6 F 15.0
28:00 260.3 F 21.0
30:00 270.2 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.3 F 28.7
2:00 272.5 F 29.1
3:00 272.1 F 29.2
4:00 271.9 F 28.6

Min 271.7 F 28.5
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.0 F 0.0
5:00 237.5 F 0.0
10:00 232.2 F 0.0
15:00 173.9 F 0.0
20:00 165.5 F 0.0
25:00 144.0 F 0.0
30:00 146.7 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 31:38
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:49
CYCLE COMPLETE

08/30/2022 08:07 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____