

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2350

Name: RITTER 1  
Sterilizer ID: V2178198  
09/21/2022 12:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 116.1 F 0.0  
2:00 131.2 F 1.6  
4:00 158.6 F 4.5  
6:00 182.2 F 8.4  
8:00 213.0 F 12.5  
10:00 217.6 F 5.1  
12:00 223.0 F 9.3  
14:00 237.7 F 11.2  
16:00 234.8 F 9.8  
18:00 240.5 F 9.2  
20:00 244.8 F 12.8  
22:00 255.4 F 18.6  
24:00 266.5 F 25.2  
26:00 272.0 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.1  
1:00 273.1 F 29.9  
2:00 272.5 F 29.3  
3:00 272.0 F 29.3  
4:00 273.0 F 29.9

Min 271.9 F 29.0  
Max 273.1 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.3 F 0.0  
5:00 200.4 F 0.0  
10:00 216.2 F 0.0  
15:00 220.3 F 0.0  
20:00 223.4 F 0.0  
25:00 229.6 F 0.0  
30:00 237.3 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 26:37  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:04:41  
CYCLE COMPLETE

09/21/2022 01:26 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_