
Midmark M11 - v1.0.5
Total Cycles: 1583

Name: RITTER2
Sterilizer ID: V2313228
10/12/2022 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.3 F 0.0
2:00 191.2 F 3.3
4:00 190.7 F 7.9
6:00 221.1 F 13.5
8:00 228.0 F 9.2
10:00 233.2 F 9.4
12:00 243.1 F 13.2
14:00 243.3 F 12.5
16:00 256.3 F 19.4
18:00 267.9 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.6 F 29.8
2:00 271.8 F 29.1
3:00 271.4 F 29.4
4:00 272.5 F 29.7

Min 271.1 F 28.8
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.1 F 0.0
5:00 220.0 F 0.0
10:00 230.9 F 0.0
15:00 225.5 F 0.0
20:00 222.1 F 0.0
25:00 220.1 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:18
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

10/12/2022 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____