
Midmark M11 - v1.0.5
Total Cycles: 2836

Name: RITTER2
Sterilizer ID: V2313228
09/06/2023 10:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.4 F 0.0
2:00 187.8 F 3.1
4:00 202.0 F 8.0
6:00 219.9 F 13.4
8:00 227.9 F 8.9
10:00 234.1 F 10.0
12:00 241.3 F 12.4
14:00 240.7 F 11.3
16:00 254.3 F 18.0
18:00 265.6 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.1 F 29.2
2:00 271.2 F 29.1
3:00 272.7 F 29.8
4:00 271.7 F 29.5

Min 270.9 F 28.7
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.1 F 0.0
5:00 242.6 F 0.0
10:00 228.3 F 0.0
15:00 223.1 F 0.0
20:00 193.0 F 0.0
25:00 194.4 F 0.0
30:00 194.7 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 19:43
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

09/06/2023 11:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____