
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001392
9:58 21/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:51
STERILIZING 5:11
136.1 C 220kPa 5:11
136.4 C 223kPa 5:41
136.4 C 222kPa 6:11
136.1 C 222kPa 6:41
136.2 C 222kPa 7:11
136.4 C 224kPa 7:41
136.4 C 223kPa 8:11
136.2 C 224kPa 8:41
136.2 C 224kPa 8:41
VENTING 8:42
AIR DRYING 9:36
CYCLE COMPLETE 17:10

Digital Signature #
DBAD8FBDB146F490
