

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3921

Name: RITTER 1  
Sterilizer ID: V2178198  
10/19/2023 08:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.4 F 0.0  
2:00 83.0 F 0.0  
4:00 111.2 F 0.0  
6:00 126.5 F 0.0  
8:00 150.0 F 0.3  
10:00 169.5 F 3.1  
12:00 203.2 F 6.6  
14:00 221.9 F 10.0  
16:00 236.8 F 13.5  
18:00 229.4 F 7.1  
20:00 237.8 F 11.2  
22:00 231.8 F 7.1  
24:00 244.9 F 13.0  
26:00 241.4 F 11.2  
28:00 254.5 F 17.4  
30:00 265.0 F 23.8  
32:00 270.9 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.5  
1:00 272.4 F 29.1  
2:00 272.6 F 28.9  
3:00 271.4 F 28.6  
4:00 272.9 F 29.3  
  
Min 271.3 F 28.4  
Max 273.2 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.4 F 0.0  
5:00 240.1 F 0.0  
10:00 236.7 F 0.0  
15:00 228.4 F 0.0  
20:00 226.7 F 0.0  
25:00 231.0 F 0.0  
30:00 230.2 F 0.0  
  
Drying Complete

FILLING: 0:48  
HEATING: 32:37  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:10:49  
CYCLE COMPLETE

10/19/2023 09:19 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_