
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001638
10:44 30/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:44
STERILIZING 3:50
136.4 C 221kPa 3:50
136.3 C 224kPa 4:20
136.3 C 223kPa 4:50
136.3 C 223kPa 5:20
136.4 C 224kPa 5:50
136.3 C 224kPa 6:20
136.3 C 222kPa 6:50
136.4 C 224kPa 7:20
136.4 C 224kPa 7:20
VENTING 7:20
AIR DRYING 8:09
CYCLE COMPLETE 15:06

Digital Signature #
B63441FC3A43F37C
